

Development and validation of a selective SPR-aptasensor for the detection of anticancer drug irinotecan in human plasma samples

Adelina Puscasu,¹ Martina Zanchetta,^{1,2} Bianca Posocco,¹ David Bunka,³ Stefano Tartaggia,^{1,*} Giuseppe Toffoli¹

¹ Clinical and Experimental Pharmacology, Centro di Riferimento Oncologico di Aviano (CRO) IRCCS, Via Franco Gallini 2, 33081, Aviano, Italy; E-mail: stefano.tartaggia@cro.it

² Department of Chemical and Pharmaceutical Sciences, University of Trieste, Via Giorgieri 1, 34127, Trieste, Italy.

³ Aptamer Group, Suite 2.78 – 2.91, Bio Centre, Innovation Way, Heslington, York, YO10 5NY, UK.

* Corresponding author. Clinical and Experimental Pharmacology, Centro di Riferimento Oncologico di Aviano (CRO) IRCCS, Via Franco Gallini 2, 33081, Aviano, Italy; phone +39 0434659816, E-mail: stefano.tartaggia@cro.it

Abstract. In this work, a surface plasmon resonance (SPR)-based assay for the quantification of anti-neoplastic drug irinotecan in human plasma samples has been developed for the first time. The selective binding of irinotecan with an aptamer receptor, operating in human plasma, allowed to set-up a novel analytical methodology to detect the drug in the analytical range of interest by using SPR as detection technique. After hybridizing the aptamer to the sensing platform and optimizing the sample preparation procedure, a quantitative assay was validated according to FDA regulatory guidelines. The analytical working range was found between 100 and 7500 ng mL⁻¹ with negligible interferences from plasma components and co-medication associated with the administration of irinotecan. The utility of the new SPR assay was confirmed by analyzing plasma samples in parallel

with LC-MS as reference technique, providing a new analytical tool for the therapeutic drug monitoring of irinotecan in patients under chemotherapy regimens.


Keywords: irinotecan, aptamers, therapeutic drug monitoring, anticancer drugs, surface plasmon resonance, oncology

Introduction

Antineoplastic drug Irinotecan (CPT-11) is a camptotecin derivative, widely indicated in the first-line treatment of patients with metastatic carcinoma of the colon or rectum [1, 2]. Chemotherapeutic regimens involving CPT-11 often include the co-administration of other anticancer drugs such as 5-fluorouracil (5-FU) and leucovorin (i.e. FOLFIRI treatment) [3, 4], or monoclonal antibodies, such as cetuximab and bevacizumab [5, 6].

Antitumor activity of CPT-11 is due to the generation of a 100-1000 fold more cytotoxic metabolite SN-38 from metabolic processes (scheme 1), which is acting as a topoisomerase I inhibitor [7, 8]. This nuclear enzyme plays a key role in relaxing the supercoils of the DNA double helix during replication, transcription and repair [9, 10]. Therefore, inhibition of this enzyme causes irreversible DNA damage inducing death of cancer cells by apoptosis.

Other main metabolites known from the metabolic profile of CPT-11 are produced through the cleavage of the (1,4'-bipiperidine) chain from the core structure of CPT-11. In particular, APC and NPC are generated by cytochrome P-450 3A4 (CYP3A4), and SN-38 is further converted into its glucuronide derivative SN-38G by hepatic UDP-glucuronyltransferase, which is excreted through urine (Scheme 1) [11].



Scheme 1 Metabolic pathway of CPT-11 with its main metabolites. CYP3A4: cytochrome P-450; UGT1A1/7: UDP-glucuronyltransferase; CE: liver carboxylase.

CPT-11 is administered *via* intravenous infusion for 2 hrs at a dosage of 350 mg/m² every three weeks in monotherapy, and 180 mg/m² every two weeks in combination with folinic acid and 5-FU [3, 4, 12]. The dose should be modified after the first administration according to the degree of adverse effects observed in patients. The side effects of irinotecan are common to most antineoplastic agents and include neutropenia, thrombocytopenia, anemia, acute cholinergic syndrome, alopecia and gastrointestinal disorders [2, 13, 14].

In recent years, the practice of therapeutic drug monitoring (TDM) has received a growing interest to personalize the drug dosage for any individual patient, but its application in oncology is still very limited [15]. In this context, point-of-care testing (PoCT) and biosensing technologies would represent a convenient strategy to extend TDM to a wide number of patients. Indeed, implementing TDM in the clinical practice would ensure that the plasma level of administered drugs is maintained within a therapeutic window of efficacy, preventing therefore most of toxic effects or treatment failures [16].

The most common methodologies utilized in centralized laboratories for the quantification of drugs in biological fluids, including CPT-11 and its main metabolites, are currently based on reverse-phase high-performance liquid chromatography (RP-HPLC) coupled with mass spectrometry [17, 18, 12]. However, the development of novel analytical tools at a PoCT level presents several technical challenges related to the analysis of therapeutic drugs, which are often low-molecular weight organic molecules, in biological fluids (blood, plasma, serum, or urine). As general considerations, new analytical tools should minimize sampling procedures and avoid or simplify pre-analytical treatments of biological samples with respect to standard chromatographic methods. Moreover, any analytical method, utilized in the clinical practice, should be validated according to regulatory guidelines, in order to provide clinicians and medical staff with reliable quantitative information to take critical decisions. Finally, portable analyzers could be used to monitor patients in clinical settings, where well-equipped laboratories are not available, such as at patients' bedside, in medical practices or, eventually, at home. This would allow testing to be carried out without the need for personnel to receive special training.

Example of sensing approaches for the detection of CPT-11 and its active metabolite SN-38 are already described in the literature, evidencing the need for the development of new analytical tools applied to anticancer drugs for TDM studies.

In particular, we previously reported an electrochemical biosensor to detect CPT-11, which was based on measuring the inhibition of acetylcholine esterase enzyme in the presence of CPT-11, with excellent selectivity toward its metabolites in simple and complex matrices [19]. An electrochemical biosensor employing aptamer receptors was also developed for *in-vivo* detection of CPT-11 and applied to animal models [20]. Instead, direct detection of CPT-11 in human samples was described by using electrochemical methods, such as adsorptive stripping square-wave voltammetry and differential pulse voltammetry. These techniques were successfully applied to the detection of CPT-11 in biological samples [21, 22, 23]. A simple and efficient fluorimetric assay to detect the active metabolite SN-38 was also validated in plasma [24].

Herein, we describe a protocol for the analysis of CPT-11 at clinically relevant concentrations by using SPR as the detection technique, and an ss-DNA aptamer as a selective receptor operating in human plasma.

Aptamers are single-stranded oligonucleotides (RNA or DNA), capable of recognizing and binding different targets such as proteins, peptides, drugs and small molecules with high affinity and specificity. They can fold into a large variety of secondary structures with specific structural motifs that form along the nucleotide chain (hairpin, pseudo-knot, stem-loop / bulge and G-quadruplex), providing a three-dimensional structure similar to those observed in proteins. Aptamers undergo adaptive conformational changes and their three-dimensional folding creates a specific binding site. In the presence of the target, the aptamer structure can be stabilized by the establishment of weak interactions [25].

Our protocol was developed using SPR, which allowed the characterization of the steady-state affinity of the aptamer-CPT11 interaction and to develop an assay for the quantification of CPT-11 in plasma samples. The working principle of SPR is typically based on the real-time monitoring of the interaction between the receptor, immobilized on the sensor surface, and analyte molecules in the samples. When a mass change (deriving from such interaction) is occurring at the sensor surface, a change of refractive index at the interface is originated. This SPR signal is recorded in response units (RU). Large biomolecular systems such as antigen-antibody are commonly investigated with SPR [26], while the interaction of low molecular weight targets (<1000 Da) with large biomolecular receptors (typically >50000 Da) results in the generation of a significantly weaker SPR signals, due to the small local variation of refractive index at the surface [27, 28]. This therefore presents a challenge for monitoring small molecule interactions with high molecular weight affinity ligands (such as antibodies). A typical SPR measurement is based on the sequential injection of the sample across a 'blank' reference flow cell (FC1) and a working 'receptor immobilized' flow cell (FC2). Information relating to the analyte-receptor binding can be therefore obtained through a differential

measurement (FC2-FC1), enabling subtraction of any interference generated by non-specific surface interactions.

Materials and methods

Chemicals

Analytical standards of CPT-11 (7-ethyl-10-[4-(1-piperidino)-1-piperidino]-carbonyloxycamptothecin, purity 97%), SN-38 (7-ethyl-10-hydroxycamptothecin, purity 98%) were purchased from Sigma-Aldrich Co. (Milan, Italy). APC (7-ethyl-10-[4-N-(5-aminopentanoic acid)-1-piperidino]-carbonyloxycamptothecin, purity 95%), NPC (7-ethyl-10-(4-amino-1-piperidino) carbonyloxycamptothecin, purity 98%) and SN-38G (7-ethyl-10-[3,4,5-trihydroxy-pyran-2-carboxylic acid]-camptothecin, purity 95%) were purchased from Toronto Research Chemicals, Inc. (North York, Ontario, Canada). PBS buffer, magnesium chloride ($MgCl_2$), acetic acid were purchased from Merck Sigma-Aldrich (Milan, Italy). Tween20 was purchased from VWR (Milan Italy). CPT-11 aptamer (100-mer), 5'biotin CPT-11 aptamer and 5'biotin immobilization oligomer (15-mer) (see details in Supporting Information and in the recently reported patent [29]) were purchased from Aptamer Group (York, UK). Lyophilized oligomers were solubilized with milli-Q water upon arrival, aliquoted and stored at $-20^{\circ}C$. Co-medications for selectivity tests, in particular dexamethasone, atropine, chlorphenamine, 5-fluorouracil, ondansetron, loperamide and capecitabine, were provided by the pharmacy of the National Cancer Institute (Aviano, Italy) and folinic acid was purchased from Merck Sigma-Aldrich. LC-MS grade methanol, acetonitrile and camptotecin were supplied by Merck Sigma-Aldrich. Control human plasma stabilized with K_2EDTA for the preparation of daily standard calibration curves and quality control (QC) samples was obtained from healthy volunteers and was provided by the Transfusion Unit of the Centro di Riferimento Oncologico di Aviano (CRO), Italy.

SPR measurements

SPR measurements were performed using a Biacore X100 workstation (GE Healthcare Life Sciences), at 25 °C, with a flow rate of 5, 10 and 30 $\mu\text{L min}^{-1}$. Streptavidin coated sensor chips (SA) were purchased from GE Healthcare Life Sciences.

Binding analysis

Binding analysis was performed by using 13 different concentrations of CPT-11 from 0 to 2.0 μM in PBS buffer (10 mM, pH 6.0) containing NaCl (100 mM), KCl (2 mM) MgCl_2 (7.5 mM) and Tween 20 (0.05%) on a SA chip functionalized with 5'biotin CPT-11 aptamer (see Supporting Information, Fig. S1 and S2). Experiments were performed in triplicate. Association and dissociation time were 120 and 180 s respectively, with the flow rate set to 30 $\mu\text{L/min}$. Regeneration of the sensor surface was performed with H_2O for 20 s after each sample injection. Data processing was performed by subtracting the signal recorded on FC1 (reference cell) from the signal recorded on FC2 (functionalized cell). FC2-FC1 data were further referenced by blank subtraction. Association curves were thus evaluated by non-linear analysis with SPR kinetic evaluation software (BIAevaluation Software, version 2.0.2 Plus Package, GE Healthcare Life Sciences) and the affinity constant was obtained from a 1:1 steady state model.

Concentration analysis in buffered media

CPT-11 aptamer (1.2 μM) in PBS buffer (50 mM, pH 7.2) containing NaCl (685 mM), KCl (13.5 mM) and MgCl_2 (5 mM), defined as hybridization buffer, was injected on FC2 of the SA chip functionalized with 5'biotin immobilization oligomer (see Supporting Information, Fig. S1 and S3) over 600s at 5 $\mu\text{L min}^{-1}$. After a stabilization time of 300 s, standard solutions of CPT-11 were sequentially injected on FC1 and FC2 over 420 s at a flow rate of 10 $\mu\text{L/min}$. Calibration standards of CPT-11 from 50 to 800 ng mL^{-1} in PBS buffer (10 mM, pH 6.0) containing NaCl (100 mM), KCl (2 mM) and MgCl_2 (2.5 mM), were used. The running buffer was PBS (25 mM, pH 7.2) containing

NaCl (342 mM), KCl (6.75 mM) and Tween20 (0.05% v/v). Surface regeneration was performed at the end of each run with a mixture of 8 mM NaOH and 160 mM NaCl for 20 s. A blank sample (0 ng/mL) was also included.

Concentration analysis in human plasma

Standards and quality control working solutions

A 1000 $\mu\text{g mL}^{-1}$ stock solution of CPT-11 was firstly prepared in DMSO, aliquoted and stored at -20 °C. A series of working solutions (A to G) were obtained by diluting the stock solution with DMSO to the concentrations reported in Table S1 (Supporting Information) and used to prepare calibration standards and quality control (QC) samples at low (QCL), medium (QCM) and high (QCH) concentrations. Aliquots of these solutions were stored at -20 °C in Eppendorf polypropylene tubes and used for a maximum of three freeze-thaw cycles.

Preparation of standards and quality control samples

A newly prepared seven-point calibration curve was recorded every day during this study. Each calibration and QC sample was prepared by adding 2 μL of working solution from A to G in 38 μL of pooled human plasma in a 0.5 mL Eppendorf polypropylene tube to obtain the final concentrations reported in Table S1 (Supporting Information). Each calibration curve included a blank and three QC samples, which were analyzed in triplicate. Calibration standards and QCs were diluted with 360 μL of PBS buffer (12 mM, pH 5.0) containing NaCl (103 mM), KCl (2 mM) MgCl_2 (2.5 mM) and Tween 20 (0.05%), named as incubation buffer, and thoroughly vortexed for 10 seconds three times. The content of each tube was transferred on a centrifugal filter (Millipore Amicon® Ultra 0.5 mL, 30 kDa cut off) and centrifuged at 6200 rcf for 18 min at 25°C. Then, 350 μL of the filtrate were transferred

on 11 mm plastic vials (Scheme S1, Supporting Information) and analyzed as described below. CPT-11 standards and samples were kept in ice bath until use.

Concentration analysis

Concentration analysis of plasma samples were performed with the SA chip functionalized with 5'biotin immobilization oligomer (Supporting Information, Fig. S1 and S3), using PBS (25 mM, pH 7.2) containing NaCl (342 mM), KCl (6.75 mM) and Tween 20 (0.05% v/v) as the running buffer. For each run, CPT-11 aptamer (1.2 μ M) in the hybridization buffer was firstly injected in FC2 for 600 s at 5 μ L/min. After 300 s of stabilization time, CPT-11 standards were injected in both flow cells at 10 μ L/min for 480 s in the following order: blank, calibration standards, QCs in triplicate and unknown samples. After each run, the chip surface was regenerated with a mixture of 8 mM NaOH and 160 mM NaCl for 20 s.

Validation study

The validation of the method was carried out by following the guidelines of the Food and Drug Administration (FDA) applied to ligand-binding assays (LBA) [30]. The parameters evaluated for the validation included recovery, intra- and inter-day precision and accuracy, reproducibility, lower limit of quantification (LLOQ), selectivity and matrix effects (see Supporting Information for full details).

Mass Spectrometry

Quantification of CPT-11 in plasma samples, collected from patients on therapy, was performed by using tandem mass spectrometry, as the reference method to assess the quality of the SPR assay. The analysis was conducted with a Shimadzu LC system coupled with an API 4000 QT (Sciex) working in multiple reaction monitoring (MRM) mode. Irinotecan was quantified in plasma samples applying a LC-MS/MS method previously developed and validated, according to FDA/EMA guideline, by our

group.[12] In brief, 100 μL of the actual sample, calibrator or QC sample were transferred to a 1.5 mL polypropylene tube and mixed with 5 μL of the internal standard solution (camptothecin, 0.5 $\mu\text{g}/\text{mL}$), and successively with 300 μL of 0.1% $\text{CH}_3\text{COOH}/\text{CH}_3\text{OH}$ for protein precipitation. The solution was vortexed and centrifuged at 16000 g for 10 min at nominally 4°C. Then 150 μL of the obtained supernatant were transferred to an autosampler glass vial for the analysis. Analytes separation was obtained with gradient elution on a Gemini C18 column (3 μM 110A, 100 x 2.0 mm) coupled with a Security Guard Cartridge (Gemini-NX C18 4.0 x 2.0 mm) (column temperature set at 25°C). The mobile phases (MP) were 0.1% $\text{CH}_3\text{COOH}/\text{double distilled water}$ (MP A) and 0.1% $\text{CH}_3\text{COOH}/\text{acetonitrile}$ (MP B). The ESI source operated in positive ion mode at 650°C and with ion spray voltage of 5500 V. Quantification was performed using the following transitions: m/z 587 > 124 for CPT-11 and m/z 349 > 305 for the internal standard.

Results and discussion

The CPT-11 aptamer was preliminary tested to establish the optimal experimental conditions to detect CPT-11. Our approach included also the utilization of 5'biotin immobilization oligomer, which is a short sequence (15-mer) complementary to a hybridization region within the CPT-11 aptamer (100-mer) [29]. The functionalization of the SPR platform was easily carried out by using 5'biotin- CPT-11 aptamer and immobilization oligomer, which were successfully immobilized on a streptavidin coated (SA) sensor chip (see Supporting Information, Fig. S1). Indeed, the simple injection of a 5 μM solution of biotinylated DNA in 25 mM PBS buffer provided an immobilization yield of 1767 and 2092 RU on FC2 for 5'biotin- CPT-11 aptamer and immobilization oligomer respectively (see Supporting Information, Fig. S2 and S3). The surface density for the aptamer and the immobilization oligomer on the chip was estimated in 1.767 and 2.092 ng mm^{-2} respectively, after conversion of the

corresponding RU as reported for Biacore technology, in which a response of 1000 RU corresponds to a change in surface density of about 1 ng mm^{-2} of proteins or DNA [31].

Steady state affinity study

The evaluation of the affinity constant (K_D) between CPT-11 aptamer and CPT-11 was carried out using an SA chip prepared by immobilizing 5'biotin CPT-11 aptamer on FC2 only, as described in Supporting Information (Fig. S2). After injecting CPT-11 in PBS buffer at a flow rate of $30 \mu\text{L}/\text{min}$ across both FC1 and FC2, the response, obtained from the differential measurement (FC2-FC1), rapidly achieved a plateau. As soon as the injection stopped, the signal decreased immediately, demonstrating that both association and dissociation of CPT-11 from the aptamer were fast processes, as shown in Fig. 1a.




Fig. 1 (a) reference subtracted sensorgrams (FC2-FC1) and (b) steady state affinity curve obtained after injecting CPT-11 in the concentration range $0 - 2.00 \mu\text{M}$ in PBS buffer (10 mM, pH 6.0) containing MgCl_2 (7.5 mM) and Tween 20 (0.05%) on the SA chip functionalized with 5'biotin CPT-11 aptamer. To evaluate K_D , binding values were collected four seconds before injection stop and were further referenced by blank subtraction. Experiments were performed in triplicate.

The affinity constant (K_D) was evaluated by analyzing CPT-11 in a concentration range from 0 to 2 μ M and fitting the steady state affinity curve, obtained by plotting the RU at the plateau against the concentration of CPT-11 for each curve, with a 1:1 binding model (Fig. 1b, BIAevaluation Software, version 2.0.2 Plus Package, GE Healthcare Life Sciences). Even though the binding of CPT-11 provided a relatively weak SPR response, the model gave a K_D of 285 ± 17 nM, a typical value for the binding of small molecules [28].

Detection of CPT-11 in buffered media

Although direct binding of CPT-11 with the aptamer immobilized on the sensor surface provided a measurable SPR response, we pursued a different strategy to develop a concentration assay. In fact, the aptamer has a molecular weight (MW) of 31000 Da, which is considerably higher than the CPT-11 molecule (MW of 586.7). Binding of the aptamer has a higher refractive index change at the interface and consequently results a higher SPR response from the binding process. Interferences from non-target molecules (e.g. plasma proteins) would hamper the analysis of real samples. For this reason, we were discouraged from follow the direct detection approach. We then moved to a strategy based on functionalizing the SA chip with 5'biotin immobilization oligomer on both flow cells (see Supporting Information, Fig. S3), followed by the hybridization of CPT-11 aptamer to form a duplex structure on FC2. Under this configuration, the conformational change of the aptamer, in the presence of CPT-11, would lead to the release of the aptamer from the duplex structure as shown in Fig. 2 [32]. As a result, the mass decrease from the sensor chip would be significantly greater and more consistent than the direct binding of CPT-11 to the aptamer immobilized at the surface.

During our investigation, we found a satisfactory hybridization level of CPT-11 aptamer on SA sensor chip by using a 1.2 μ M solution of the aptamer in 50 mM PBS containing 5 mM $MgCl_2$ (hybridization buffer) over the sensor surface for 600 s at 5 μ L/min. A PBS concentration of 50 mM was beneficial to increase the hybridization level, even though higher salt concentrations were avoided to prevent

the possible formation of salt precipitates in the microfluidic cell of the instrument. Moreover, PBS concentration was kept at 25 mM in the running buffer to reduce the non-specific displacement of the aptamer from the surface. On the other hand, the surface regeneration could be efficiently performed with a solution of 8 mM NaOH and 160 mM NaCl for 20 sec, thus enabling the possibility to recycle the SA chip after each experiment.




Fig. 2 Schematic representation of the SPR aptasensor to detect CPT-11. After hybridization of the CPT-11 aptamer on the chip surface in FC2, the injection of a sample containing CPT-11 lead to the release of the aptamer from the chip surface with the generation of a SPR response suitable for analytical quantification.

According to our tests, the aptamer hybridization level decreased from 1100 to 1000 RU after 400 runs with an average value of 1053 ± 34 RU and a CV of 3.2% (see Supporting Information, Fig. S4). After such usage, a loss of activity occurred due to an excessive degradation of the surface and the chip was replaced. The aptamer was hybridized only on FC2, while FC1 remained unfunctionalized and was used as reference cell, to exclude non-specific interactions of CPT-11 with the sensor surface during data processing.

Upon hybridization of the aptamer on FC2, solutions of CPT-11 were sequentially injected in both flow cells. Dissociation of aptamer from FC2 occurred with a decrease of SPR signal proportional to the concentration of CPT-11 (Fig. 3a). Several buffer compositions were tested in order to optimize the interaction between the aptamer and CPT-11, providing that 10 mM PBS at pH 6 containing

MgCl₂ in the range 2.5 – 7.5 mM could be used for the incubation. After the incubation step, the chip was regenerated for the next run.

Sensorgrams were processed in two steps as follows. (i) SPR data of each run were referenced by subtracting the response of FC1 (reference cell) from FC2 to provide FC2-FC1 data (Fig. 3a), and (ii) sensorgrams were further corrected by subtracting the blank from each non-zero sample and ‘flipping of data’ to plot the aptamer dissociation as a positive response (Fig. 3b).




Fig. 3 (a) Significant portions of sensorgrams for the overall calibration experiment (FC2-FC1). The three sections are summarized as follows: (i) injection and hybridization of the aptamer on the sensor chip surface, (ii) sample injection and (iii) regeneration of the chip surface. (b) Flipped sensorgrams obtained after blank subtraction, to plot the aptamer dissociation as a positive response. (c) Calibration curves obtained by plotting Δ RU values, evaluated 200 s after the end of sample injection, vs CPT-11 concentrations. Mg = MgCl₂. Experiments were performed in triplicate.

The correlation of Δ RU values, which were taken 200 s after the stop of each sample injection, against the concentration of CPT-11, provided a response in the concentration range from 50 to 800 ng mL⁻¹

that could be fitted by using a non-linear regression (Fig. 3c). A four parameter logistic (4PL) model was applied following FDA guidelines for LBA validation (see Supporting Information for a brief description of the model) [30]. We put a huge effort to obtain a linear response during this phase of the method development, including the modification of the buffer composition as well as the utilization of different contact times and flow rates. Unfortunately, any variation of experimental conditions had an impact on the relative intensity of the SPR response but not in the shape of the calibration curve. In general, more diluted buffers led to a higher response, whereas media with higher ionic strength reduced the sensitivity of the method, due to an improved stability of the aptamer-hybridized to the surface. As an example, the two calibration curves shown in Fig. 3c, obtained diluting CPT-11 with a PBS buffer (10 mM) containing 2.5 and 7.5 mM MgCl₂ respectively, clearly showed a higher response when using a more diluted buffer. The detection limit in buffered media was evaluated by analyzing the blank variability (three times the standard deviation of the blank on eight replicates) with the 4PL model, providing a LOD of 9.1 ng mL⁻¹.

Detection of CPT-11 in human plasma

Having found a strategy suitable to detect and calibrate CPT-11, we next moved to the application of our method to human plasma samples. Contrary to a simple buffered media; human plasma is a complex matrix with a high protein content, especially albumin and α -glycoproteins, which are known to bind 60-70% of CPT-11 [33].

The first step in our investigation was therefore the treatment of plasma samples for SPR analysis. Direct injection of plasma samples into the flow cells of Biacore X100 instrument was avoided to reduce the risk of blockages to the instruments microfluidics.

A 1:10 dilution step was applied to match the instrumental response, previously obtained between 50 and 800 ng mL⁻¹, with a concentration range of CPT-11 that can be reasonably found in plasma samples, usually up to 5000 ng mL⁻¹ [12].

Unfortunately, the simple dilution of plasma was not sufficient to allow SPR analysis, as plasma proteins still contributed a significant response due to adsorption on the sensor surface, obscuring the aptamer dissociation process. An additional step of microfiltration with 30 kDa centrifugal filters before SPR analysis was introduced to completely avoid the response of large protein content contained in plasma. In this case, sensorgrams corresponding to the analysis of plasma appeared similar to those previously recorded in simple buffer (Fig. 3a and 3b). However, the response of CPT-11 was too weak for analytical purposes, due to the binding with plasma proteins, which prevented enough recovery of the drug in the filtrate. The addition of Tween 20 in the buffer successfully reduced non-specific binding (NSB) and allowed a significant fraction of CPT-11 to pass the centrifugal filter. Other surfactants (e.g. sodium dodecyl sulfate or cetyl trimethyl ammonium bromide) or standard dextran NSB reducer [34] (GE Healthcare Life Sciences) did not provide significant recovery from plasma. Tween 80 or PEG were also tested and found to be less effective than Tween 20 (see Supporting Information, Fig. S5). The amount of surfactant in the buffer was 0.05%.

Concentration of $MgCl_2$ in the incubation buffer was kept at 2.5 mM to have a good sensitivity, analogously to the measurement in simple buffered media. As previously observed, the correlation of SPR response against concentration of CPT-11 in plasma required a 4PL model to properly fit the calibration points, as suggested also by LBA validation guidelines (see Supporting Information for a brief description of the model) [30].

Next, once the procedure to prepare plasma samples for SPR analysis was defined, the aptamer-based assay for the quantification of irinotecan assay was validated according FDA guidelines to assess the suitability of the method.

Calibration curves

Accuracy and precision for each standard point were evaluated on six calibration curves recorded in different working days, as summarized in Table 1. The SPR shift against the nominal concentration

of the analyte was used to obtain calibration curves by using a 4PL regression (Fig. 4a). Acceptable back-calculated concentrations in a concentration range from 100 to 7500 ng mL⁻¹ were obtained with a Pearson's coefficient of determination (R²) higher than 0.995 for each set of measurement, while the accuracy was found between 94.5% and 105.2% and precision from 1.6% to 5.8% in terms of CV%.

Table 1 Summary of validation data for the SPR analysis of CPT-11. N. C. = nominal concentration, B. C. = back calculated concentration, SD = standard deviation, CV = coefficient of variation.

Entry	N. C. (ng mL ⁻¹)	Mean B. C. ± SD (ng mL ⁻¹)	Precision (%)	Accuracy (%)
calibration curves	100	105 ± 6	5.8	105.2
	250	236 ± 8	3.4	94.5
	500	508 ± 26	5.1	101.5
	1000	1004 ± 26	2.6	100.4
	2500	2504 ± 62	2.5	100.2
	5000	4977 ± 153	3.1	99.5
	7500	7521 ± 118	1.6	100.3
Intra-day	200	200 ± 7	3.5	100.2
	1500	1468 ± 6	0.4	97.8
	6500	6432 ± 45	0.7	98.9
Inter-day	200	199 ± 19	9.5	99.7
	1500	1541 ± 71	4.6	102.7
	6500	6489 ± 492	7.6	99.8
Matrix effect	100	103 ± 10	9.5	102.9
	200	181 ± 9	5.1	90.3
selectivity	1500	1470 ± 26	1.8	98
	6500	6216 ± 138	2.2	95.6
		Recovery (%) ± SD	CV %	
Recovery	200	65.0 ± 6.1	9.4	
	1500	64.3 ± 4.1	6.4	
	6500	73.7 ± 3.0	4.0	
		LOD = 55 ng mL ⁻¹	LLOQ = 100 ng mL ⁻¹	

Recovery

As previously described, the preparation of samples for the SPR analysis was carried out by diluting plasma 1:10 with an incubation buffer followed by centrifugal microfiltration with 30 kDa cut-off centrifugal filters (Supporting Information, Scheme S1). To evaluate the recovery of CPT-11 from the sample preparation procedure, the SPR response of QC samples, prepared by spiking CPT-11 in plasma before the dilution and centrifugal microfiltration, was compared to the response of QCs in which CPT-11 was added after the dilution and microfiltration of plasma. Recovery was evaluated at three concentration levels and was found in the range 64.3 – 73.7% with $CV \leq 10\%$ (Table 1). The significance of recovery, in our method, is referred to the efficiency of centrifugal microfiltration procedure to recover the analyte from the sample, and indicates the fraction of CPT-11 found in the filtrate, with respect to the theoretical amount present in the sample. Even though the microfiltration step was found to result in a significant loss of the analyte, it must be considered that the same procedure was applied to both calibration standards and samples. In fact, standards and QCs were prepared by spiking working solutions of CPT-11 in plasma before the microfiltration, to simulate the treatment of real samples as much as possible. To confirm this hypothesis, we prepared QCs using independent stock solutions of CPT-11 than those utilized for calibration standards, and determined intra- and inter-day accuracy and precision, as described below.

Intra-day and inter-day precision and accuracy and reproducibility

To establish the precision and accuracy of the method, we analyzed QC samples in triplicate (200, 1500, 6500 ng mL⁻¹) within a single run for intra-day and over six sets of measurement for the inter-day assessment. The intra- and inter-day accuracy was found in the range 97.8 – 100.2% and 99.7-102.7% respectively. Moreover, the method demonstrated an excellent intra- and inter-day precision (CV%) in the range 0.4-3.5% and 4.6-9.5% (Table 1). Inter-day precision and accuracy were also evaluated by combining data collected from three different SA chips at the QC concentration levels, in order to assess the reproducibility of the method using different chips. In particular, we found a

precision (CV%) of 5.7, 2.9 and 6.2% at 200, 1500 and 6500 ng mL⁻¹, whereas accuracy at the same concentrations was 99.4, 96.6 and 99.4% respectively.

Lower limit of quantification (LLOQ), limit of detection (LOD) and matrix effect

According to guidelines by FDA, a bias up to 25% is allowed at the LLOQ level for LBA tests, even though a tighter value was suggested to be beneficial when fitting a model to the standards [35]. Having used a non-linear fitting for the preparation of the calibration curve, we evaluated LLOQ by analyzing the biases of eight independent replicates for each standard, which were plotted against the nominal concentration of CPT-11. Biases were within 15% for all calibrators except at 100 ng mL⁻¹, in which a $\pm 17\%$ value was found appropriate as the LLOQ and the lowest standard of the calibration curve (see Supporting Information, Fig. S6). Lower concentrations were not considered, as the absolute SPR response at 100 ng mL⁻¹ was found between 7.5 and 10.4 RU, which is, reasonably, the lowest response for quantitative detection of CPT-11. Therefore, lower responses would have been inaccurate. The LOD of the method was evaluated on eight replicates of the blank within a single set analysis. The analysis of the blank variability (three times the standard deviation of the blank), with the 4PL model, resulted a LOD of 55 ng mL⁻¹ in plasma. Consequently, this method can be applied to evaluate irinotecan in samples taken from patients up to 20 hrs after intravenous infusion, in which the concentration of CPT-11 is close to 100 ng mL⁻¹ [12]. The method was also designed to detect high concentrations of the drug, which are leading to excessive toxicity, enabling therefore to properly adjust the administered dosage.

The evaluation of matrix effects and other interferences in the sample was tested at the LLOQ with six sources of human plasma, obtained from different donors. The quantification through a calibration curve, prepared with pooled plasma, provided the same back calculated concentration at 100 ng mL⁻¹ with a precision of 9.5% and accuracy of 103.0%. These values demonstrate that the microfiltration of samples to remove most of the biological interference is effective, whereas the aptamer maintain

the selectivity toward the recognition of CPT-11 against those low molecular-weight species remaining in the sample.

Selectivity

Selectivity was evaluated by testing the interaction with other known interfering low-MW compounds found in samples, which included co-medications such as dexamethasone, atropine, chlorphenamine, 5-fluorouracil, ondansetron, loperamide and capecitabine and folic acid, commonly associated with the administration of CPT-11, and the main metabolites of irinotecan SN-38, SN-38G, APC and NPC. As expected, co-medications provided a SPR response similar to that of a blank sample, demonstrating that no interaction takes place with the immobilized aptamer (see Supporting Information, Table S2 and Fig. S7). Other camptotecins spiked in plasma in pure form provided an average response of 2% for SN-38G, 11% for SN-38, 97% for NPC and 29% for APC with respect to CPT-11 at the same concentrations of QCs (See Supporting Information, Table S3). However, when analyzing QC samples containing a mixture of CPT-11 and the other metabolites, as found in real samples, we found a negligible interference with respect to the response of QCs containing CPT-11 alone. This was demonstrated from the quantification of CPT-11 in QCs containing also a mixture of metabolites having a composition of 8% APC, 4% SN38G, 1.5% SN38 and 1.5% NPC with respect to CPT-11, averaging those found in real samples (see Supporting Information, Fig. S8) [12, 34]. The quantification of CPT-11 in the presence of metabolites provided in general a negative bias with accuracies ranging from 90.3 to 98 % (Table 1), which is fully acceptable according to guidelines (bias within 20%). Interestingly, the presence of metabolites decreased the efficiency of the aptamer to detect CPT-11 and, in our opinion, this was ascribed to the ability of metabolites to inhibit the aptamer, which is prevented to freely interact with CPT-11. However, we can conclude that validation data confirmed the utility of the biosensing approach to the detection of CPT-11 in human plasma.




Fig. 4 (a) Calibration curve of CPT-11 in plasma obtained upon blank subtraction and 4PL regression. Concentration of CPT-11 corresponds to the concentration in plasma (C_{in}) reported in Table S1 (Supporting Information). Experiments were performed in triplicate; (b) correlation of SPR and HPLC-MS data for the quantification of CPT-11 in plasma samples; (c) Bland-Altman plot showing biases vs HPLC-MS.

Comparison of the SPR assay with HPLC-MS

The final test to assess the potential of our SPR assay for TDM application was the re-quantification of patient's samples previously analyzed by standard HPLC-MS reference technique. Our protocol was applied to 72 available samples, taken from patients administered with irinotecan at the *Centro di Riferimento Oncologico di Aviano (CRO)*, Italy. The samples were processed following the same procedure for standards and QCs (Supporting Information, Scheme S1). As shown in Fig 4b, the correlation of the SPR assay with respect to HPLC-MS provided a slope of 1.09 and a Pearson coefficient of 0.945. The mean difference of the values obtained from the two techniques was 0.035%,

with 54 samples out of 72 (75%) analyzed with the SPR aptasensor which displayed a deviation from the reference method lower than 20% (see Supporting Information Table S4). Moreover, an acceptable degree of correlation between the two techniques was obtained by analyzing data with the Bland-Altman plot (Fig. 4c).

Analytical performances of our methodology, in terms of analytical range and LOD, were compared to the recently reported (in the last five years) electrochemical and biosensing strategies for detecting CPT-11 in biological samples (Table 2).

Table 2 Comparison of the analytical performances of recently reported methods for detecting CPT-11 in biological samples.

Method	Analytical range (ng mL ⁻¹)	LOD (ng mL ⁻¹)	Sample type	Pre-treatment procedure (sample volume, μ L)	Reference
HPLC-MS	10 - 10000	0.116	Plasma (human)	1:4 dilution with 0.1% CH ₃ COOH/CH ₃ OH followed by centrifugation (100)	12
Pencil graphite electrodes coupled with square wave cathodic adsorptive stripping voltammetry	495 - 3639	13.1	Serum (human)	1:3 dilution with CH ₃ OH followed by centrifugation (450)	21
Solid phase extraction coupled with differential pulse voltammetry	312 - 5608	79.8	Plasma (human)	Loading on SPE columns followed by elution with CH ₃ CN (125)	23
Amperometric enzyme- based sensor	10 - 10000	1.5	Fetal bovine serum	No sample treatment (n/a)	19
Graphene quantum dots- polyaniline/zinc oxide nanocomposites electrochemical sensor	62 - 15579	8.7	Serum (human)	1:50 dilution with Britton- Robinson buffer (n/a)	22
Electrochemical aptamer- based sensor	60 - 10000	n/a	Whole blood (rat)	In-vivo measurement after surgery (n/a)	20
SPR aptamer-based sensor	100 - 7500	55	Plasma (human)	1:10 dilution with incubation buffer followed by centrifugal microfiltration (40)	This work

Although LOD is still far from the reference HPLC-MS method (0.116 ng mL⁻¹) [12], our SPR aptasensor displayed a working range (100 – 7500 ng mL⁻¹) and a LOD (55 ng mL⁻¹) of the same

order of magnitude as reported for other detection approaches. The main advantage of the SPR aptasensor with respect to chromatographic [12] and electrochemical [21][23] methodologies is that no organic solvents are needed for the analysis, making our approach more suitable to be transferred at a PoCT level. Moreover, our protocol for CPT-11 detection was applied to human plasma, instead of simpler matrix, such as fetal bovine serum utilized to test our previously reported amperometric enzyme-based sensor [19]. Although the analytical range and the LOD of our method do not achieve better values than direct electrochemical [22] and electrochemical aptamer-based [20] sensors, we developed our methodology to exclude interferences, especially by metabolites and co-medications of CPT-11, and finally performed a cross-validation with HPLC-MS, to verify the utility of our method. On the other hand, the two main limitations are represented by the sample pre-treatment and the time to analysis. In particular, the centrifugal microfiltration of plasma limited the recovery of the drug from the sample to 64.3 – 73.7% and was found a time consuming step. Overall time to process and analyze samples, including the run time for SPR analysis, was around 1 hour. Ideally, this is needing to be further shortened, in order to respond to the timely need for drug administration or reduction.

Considering the excellent validation and cross-validation data of the method, obtained from this pilot study, we conclude that a biosensing approach employing aptamers would be beneficial to the development of novel analytical tools for the quantification of irinotecan in clinical samples. This is a further example showing that advances in biosensor technology would extend the practice of TDM in oncology to optimize the drug dosage during administration and prevent those toxicity events occurring many hours or days after the administration of CPT-11, especially leuco-neutropenia and diarrhea [11].

Conclusions

In conclusion, we described the development of an SPR assay to detect the anti-cancer drug irinotecan in human plasma samples employing an aptamer as a selective biomolecular receptor. Although direct

detection of small molecules with SPR provides typically a low response, we successfully developed a quantitative assay for irinotecan with consistent validation data covering the plasmatic concentrations found in clinical samples. In our opinion, a significant technical result of this study is that a simple procedure with inexpensive and non-toxic reagents can be used to prepare samples for the analysis. Furthermore, this work demonstrates that biosensor-based strategies can be implemented to the detection of small drug molecules in biological fluids with reliable validation data required for clinical analysis by regulatory guidelines. We therefore prospect that aptamer-based assays will find more space in the development of novel analytical tools aimed at improving the practice of TDM in clinical oncology.

Declarations

Conflicts of interests. D. B. is a Co-founder and Director of Aptamer Group and therefore has financial interest in Aptamer Group Ltd. A. P., M. Z., B. P., S. T. and G. T. have no conflicts to declare regarding the publication of this manuscript.

Ethics approval. Plasma samples were collected from patients enrolled in a phase Ib clinical study (CRO-2009-25, Prot. n. 0041793(09)-PRE.21-984) approved by the ethics committee of CRO-National Cancer Institute of Aviano (Italy) and by Istituto Superiore di Sanità (ISS, Italy), which was conducted according to the principles expressed in the Declaration of Helsinki. Informed consent was obtained from all participants involved in the study.

Source of biological material. Control human plasma stabilized with K₂EDTA for the preparation of daily standard calibration curves and quality control (QC) samples was obtained from healthy volunteers and was provided by the Transfusion Unit of the Centro di Riferimento Oncologico di Aviano (CRO), Italy.

Funding. This work was supported by the Italian Ministry of Health (Ricerca Corrente). We acknowledge Regione Friuli-Venezia-Giulia through the POR FESR 2014–2020 (project “Nano Diagnostics and Automated Tools for Oncology”, NADIATools) and the grant AIRC 5X1000 (Rif. 12214) “Application of advanced nanotechnology in the development of innovative cancer diagnostics tools” for financial support.

References

1. Conti JA, Kemeny NE, Saltz LB, Huang Y, Tong WP, Chou TC, Sun M, Pulliam S, Gonzalez C Irinotecan is an active agent in untreated patients with metastatic colorectal cancer. *J Clin Oncol* 1996;14:709–15.
2. Rothenberg ML, Eckardt JR, Kuhn JG, Burris HA, Nelson J, Hilsenbeck SG, Rodriguez GI, Thurman AM, Smith LS, Eckhardt SG, Weiss GR, Elfring GL, Rinaldi DA, Schaaf LJ, Von Hoff DD Phase II trial of irinotecan in patients with progressive or rapidly recurrent colorectal cancer. *J Clin Oncol* 1996;14:1128–35.
3. Tournigand C, André T, Achille E, Lledo G, Flesh M, Mery-Mignard D, Quinaux E, Couteau C, Buyse M, Ganem G, Landi B, Colin P, Louvet C, de Gramont A FOLFIRI followed by FOLFOX6 or the reverse sequence in advanced colorectal cancer: a randomized GERCOR study. *J Clin Oncol* 2004;22:229–37.
4. Guo Y, Shi M, Shen X, Yang C, Yang L, Zhang J Capecitabine plus irinotecan versus 5-FU/leucovorin plus irinotecan in the treatment of colorectal cancer: a meta-analysis. *Clin Colorectal Cancer* 2014;13:110–8.
5. Chen K, Gong Y, Zhang Q, Shen Y, Zhou T Efficacy and safety of addition of bevacizumab to FOLFIRI or irinotecan/bolus 5-FU/LV (IFL) in patients with metastatic colorectal cancer:

A meta-analysis. *Medicine* (Baltimore) 2016;95:e5221.
<https://doi.org/10.1097/MD.0000000000005221>.

6. Kotaka M, Xu R, Muro K, Park YS, Morita S, Iwasa S, Uetake H, Nishina T, Nozawa H, Matsumoto H, Yamazaki K, Han S-W, Wang W, Ahn JB, Deng Y, Cho S-H, Ba Y, Lee K-W, Zhang T, Satoh T, Buyse ME, Ryoo B-Y, Shen L, Sakamoto J, Kim TW Study protocol of the Asian XELIRI Project (AXEPT): a multinational, randomized, non-inferiority, phase III trial of second-line chemotherapy for metastatic colorectal cancer, comparing the efficacy and safety of XELIRI with or without bevacizumab versus . *Chin J Cancer* 2016;35:102.
<https://doi.org/10.1186/s40880-016-0166-3>.
7. Ohtsuka K, Inoue S, Kameyama M, Kanetoshi A, Fujimoto T, Takaoka K, Araya Y, Shida A Intracellular conversion of irinotecan to its active form, SN-38, by native carboxylesterase in human non-small cell lung cancer. *Lung Cancer* 2003;41:187–98.
8. Martino E, Della Volpe S, Terribile E, Benetti E, Sakaj M, Centamore A, Sala A, Collina S The long story of camptothecin: From traditional medicine to drugs. *Bioorganic Med. Chem. Lett.* 2017;27:701–7.
9. Liu LF DNA topoisomerase poisons as antitumor drugs. *Annu Rev Biochem* 1989;58:351–75.
10. Hsiang YH, Liu LF, Wall ME, Wani MC, Nicholas AW, Manikumai G, Kirschenbaum S, Silber R, Potmesil M DNA Topoisomerase I-mediated DNA Cleavage and Cytotoxicity of Camptothecin Analogues. *Cancer Res* 1989;49:4385–89.
11. Mathijssen RHJ, Van Alphen RJ, Verweij J, Loos WJ, Nooter K, Stoter G, Sparreboom A Clinical pharmacokinetics and metabolism of irinotecan (CPT-11). *Clin Cancer Res.* 2001;7:2182–2194.
12. Marangon E, Posocco B, Mazzega E, Toffoli G Development and validation of a high-performance liquid chromatography-tandem mass spectrometry method for the simultaneous determination of Irinotecan and its main metabolites in human plasma and its application in a clinical pharmacokinetic study. *PLoS One* 2015;10:1-18 .

<https://doi.org/10.1371/journal.pone.0118194>.

13. Canal P, Gay C, Dezeuze A, Douillard JY, Bugat R, Brunet R, Adenis A, Herait P, Lokiec F, Mathieu-Boue A Pharmacokinetics and pharmacodynamics of irinotecan during a phase II clinical trial in colorectal cancer. Pharmacology and Molecular Mechanisms Group of the European Organization for Research and Treatment of Cancer. *J Clin Oncol* 1996;14:2688–95.
14. Chabot GG Clinical Pharmacokinetics of Irinotecan. *Clin Pharmacokinet* 1997;33:245–59.
15. Widmer N, Bardin C, Chatelut E, Paci A, Beijnen J, Levêque D, Veal G, Astier A Review of therapeutic drug monitoring of anticancer drugs part two - Targeted therapies. *Eur J Cancer* 2014;50:2020–2036.
16. Meneghello A, Tartaggia S, Alvau MD, Polo F, Toffoli G Biosensing Technologies for Therapeutic Drug Monitoring. *Curr Med Chem* 2018;25:4354–77.
17. Bardin S, Guo W, Johnson JL, Khan S, Ahmad A, Duggan JX, Ayoub J, Ahmad I Liquid chromatographic-tandem mass spectrometric assay for the simultaneous quantification of Camptosar® and its metabolite SN-38 in mouse plasma and tissues. *J Chromatogr A*. 2005: 249–55.
18. D'Esposito F, Tattam BN, Ramzan I, Murray M A liquid chromatography/electrospray ionization mass spectrometry (LC-MS/MS) assay for the determination of irinotecan (CPT-11) and its two major metabolites in human liver microsomal incubations and human plasma samples. *J Chromatogr B Anal Technol Biomed Life Sci* 2008;875:522–30.
19. Alvau MD, Tartaggia S, Meneghello A, Casetta B, Calia G, Serra PA, Polo F, Toffoli G Enzyme-Based Electrochemical Biosensor for Therapeutic Drug Monitoring of Anticancer Drug Irinotecan. *Anal Chem* 2018;90:6012–19.
20. Idili A, Arroyo-Currás N, Ploense KL, Csordas AT, Kuwahara M, Kippin TE, Plaxco KW Seconds-resolved pharmacokinetic measurements of the chemotherapeutic irinotecan: In situ in the living body. *Chem Sci* 2019;10:8164–70.

21. Temerk YM, Ibrahim H, Schuhmann W Square Wave Cathodic Adsorptive Stripping Voltammetric Determination of the Anticancer Drugs Flutamide and Irinotecan in Biological Fluids Using Renewable Pencil Graphite Electrodes. *Electroanalysis* 2016;28:372–79.
22. Hatamluyi B, Es'haghi Z, Modarres Zahed F, Darroudi M A novel electrochemical sensor based on GQDs-PANI/ZnO-NCs modified glassy carbon electrode for simultaneous determination of Irinotecan and 5-Fluorouracil in biological samples. *Sensors Actuators, B Chem* 2019;286:540–49.
23. Bonazza G, Tartaglia S, Toffoli G, Polo F, Daniele S A fast method for the detection of irinotecan in plasma samples by combining solid phase extraction and differential pulse voltammetry. *Anal Bioanal Chem* 2020;412:1585–95.
24. Tartaglia S, Alvau MD, Meneghello A, Casetta B, Polo F, Toffoli G Practical fluorimetric assay for the detection of anticancer drug SN-38 in human plasma. *J Pharm Biomed Anal* 2018;159:73–81.
25. Dunn MR, Jimenez RM, Chaput JC Analysis of aptamer discovery and technology. *Nat Rev Chem* 2017;1:0076. <https://doi.org/10.1038/s41570-017-0076>.
26. Hearty S, Leonard P, Ma H, O’Kennedy R Measuring Antibody-Antigen Binding Kinetics Using Surface Plasmon Resonance. In: Nevoltris D., Chames P. (eds) *Antibody Engineering. Methods in Molecular Biology*. Humana Press, New York, NY. 2018;1827:421–55. https://doi.org/10.1007/978-1-4939-8648-4_22.
27. Cannon MJ, Papalia GA, Navratilova I, Fisher RJ, Roberts LR, Worthy KM, Stephen AG, Marchesini GR, Collins EJ, Casper D, Qiu H, Satpaev D, Liparoto SF, Rice DA, Gorshkova II, Darling RJ, Bennett DB, Sekar M, Hommema E, Liang AM, Day ES, Inman J, Karlicek SM, Ullrich SJ, Hodges D, Chu T, Sullivan E, Simpson J, Rafique A, Luginbühl B, Westin SN, Bynum M, Cachia P, Li YJ, Kao D, Neurauter A, Wong M, Swanson M, Myszka DG Comparative analyses of a small molecule/enzyme interaction by multiple users of Biacore technology. *Anal Biochem* 2004;330:98–113.

28. Chang AL, McKeague M, Liang JC, Smolke CD Kinetic and equilibrium binding characterization of aptamers to small molecules using a label-free, sensitive, and scalable platform. *Anal Chem* 2014;86:3273–78.
29. Bialas K, Reinemann C, Barnes E, Bunka D, Tolley A Aptamer Against Irinotecan. Patent Application WO2020128421 A1, 2020.
30. Food and Drug Administration, in: Guidance for Industry Bioanalytical Method Validation draft guidance, 2013. <http://www.fda.gov/downloads/Drugs/GuidanceComplianceRegulatoryInformation/Guidances/UCM368107.pdf>.
31. Liu Y, Wilson WD Quantitative Analysis of Small Molecule–Nucleic Acid Interactions with a Biosensor Surface and Surface Plasmon Resonance Detection. In: Fox K. (eds) Drug-DNA Interaction Protocols. *Methods in Molecular Biology (Methods and Protocols)*, Humana Press. 2010;613:1-23. https://doi.org/10.1007/978-1-60327-418-0_1.
32. Stoltenburg R, Nikolaus N, Strehlitz B Capture-SELEX: Selection of DNA Aptamers for Aminoglycoside Antibiotics. *J Anal Methods Chem* 2012;415697:1–14. <https://doi.org/10.1155/2012/415697>.
33. Combes O, Barré J, Duché JC, Vernillet L, Archimbaud Y, Marietta MP, Tillement JP, Urien S In vitro binding and partitioning of irinotecan (CPT-11) and its metabolite, SN-38, in human blood. *Invest New Drugs* 2000;18:1–5. <https://doi.org/10.1023/A:1006379730137>.
34. Situ C, Wylie ARG, Douglas A, Elliott CT Reduction of severe bovine serum associated matrix effects on carboxymethylated dextran coated biosensor surfaces. *Talanta* 2008;76:832–6.
35. Azadeh M, Gorovits B, Kamerud J, MacMannis S, Safavi A, Sailstad J, Sondag P Calibration Curves in Quantitative Ligand Binding Assays: Recommendations and Best Practices for Preparation, Design, and Editing of Calibration Curves. *AAPS J* 2018;20:22. <https://doi.org/10.1208/s12248-017-0159-4>.
36. Corona G, Elia C, Casetta B, Toffoli G Fast Liquid Chromatography-Tandem Mass

Spectrometry Method for Routine Assessment of Irinotecan Metabolic Phenotype. *Ther Drug Monit* 2010;32:638–46.